

Revision Record

Date	Version	Revision
3/28/2024	1.1.5.1	<ol style="list-style-type: none">1. Cursor: Supported more cursors, and more flexible cursor combination2. Fixed several bugs
3/7/2024	1.1.4.2	<ol style="list-style-type: none">1. Supported new probes: Siglent's SCP5030, SCP5030A, SCP5150, SCP55002. Measure: Optimized UX of Track and Trend; Supported to export measure data3. Fixed several bugs<ol style="list-style-type: none">a) Slow response after pressing Math buttonb) AWG: SCPI cannot set Loadc) Some bugs of USB2.0 compliance testd) Some bugs of 100Base-TX and 1000Base-T compliance test4. Some bugs on Options' unlock
	1.1.2.0	1 st release

Version Compatibility


Source Version	Object Version	Compatibility
1.1.2.0	1.1.5.1	Tested
1.1.2.0	1.1.4.2	Tested

Upgrade Instructions


Upgrade from a U-disk (USB Memory device)

WARNING: DO NOT shut off the instrument until the update is completed.

1. Copy the update file (*.ads) to a FLASH type U-disk, and then insert the U-disk into one of the USB 3.0 host ports of the instrument.
2. Click "*Utility*-> *menu*" on the top menu bar, and then click "*Maintenance* -> *Upgrade*" on the right menu bar. The following the menu should pop up and allow you to select the upgrade file




3. Click *Browse* in the menu above, and then select the correct update file (*.ads) in the pop-up resource manager




4. Click the recall icon  in the interface above and return to the upgrade dialog.


Click *Upgrade* to perform the upgrade operation:



5. The system will first copy and verify the upgrade package. After the upgrade package is validated, the following interface will appear. Click *Reboot* to continue the upgrade, or click *Cancel* to cancel it.



6. After the instrument reboots, check the version number through the *Utility-> menu* -> *System Info* to confirm if the upgrade is successful.



WARNING: DO NOT shut off the instrument until the update is completed.

Upgrade from the Web Server

A built-in web server provides an approach to control the instrument by web browser. This process doesn't require any additional software to be installed on the controlling computer. Set the LAN port correctly (see the User Manual for details), input the IP address of the instrument in the browser address bar, and then the user can browse and control the instrument on the web.



WARNING: DO NOT shut off the instrument until the update is completed.

1. Click the “FirmwareUpdate” button in the web interface



2. Select the correct update file (*.ads) stored on the computer. The instrument will automatically download the update file and perform the upgrade once the file is specified.

WARNING: DO NOT shut off the instrument until the update is completed.